



BREATH ANALYZERS

Mangal, ST-800	Professional Breath Alcohol Tester
	Digital display with red backlight Range: 0.000-0.199% BAC or 0.000-1.990% BAC / 0.000-0.995mg/L Clock Function Auto Power off Use Mouthpiece
Mangal, ST-2000	Professional Breath Alcohol Tester
	Digital display with red Backlight Range: 0.000-0.199% BAC or 0.000-0.995mg/L Use Mouthpiece Use 2xAAA alkaline batteries Operating temperature range: 5 ~ 40 degree C Warm-Up Time: within 10 seconds, Response Time: Within 5 seconds.
Mangal, AT-2009	Professional Breath Alcohol Tester
	MP-3 high precision sensor 3 Test Unit (%BAC, %BAC, mg/L) Real time, alarm time, count down timer Use Mouthpiece Use 3xAAA alkaline batteries Warning Signal: Beep 9 Volt Battery LCD Display
Mangal, AT-2017	Professional Breath Alcohol Tester
	Smart microprocessor controlled. Gives out audible sound alarm above the limit 5 Testing data memory Can be used with / without the mouthpiece supplie Battery saving design, low voltage indication. 9 Volt Battery Accuracy: 0.005% at 0.05% BAC
Mangal, AT-1100	Professional Breath Alcohol Tester
	Sensor: Advanced flat surfaced alcohol sensor Warm-up time: Within 10 seconeds Respond time: Within 5 seconds Blue-colored backlight Quick response and resume Auto power off Range: 0-0.199% BAC Warm-Up Time: within 10 seconds
Mangal, AT-6000	Professional Alcohol Breath Tester
	Advanced flat surfaced alcohol sensor SMD assembling, stable performance Direct testing process LCD indication Audio warning beyond pre-set limit Digital LCD display with light blue backup Detection Range: 0.00-0.20 BAC%; 0.00-1.00mg/L; 0.00-200mg/100mL Accuracy: ±0.015% BAC at 0.05% BAC Alarming Level: 0.05 BAC%; 0.25mg/L; 50mg/100mL
Mangal, AlcoTest-1	Professional Breath Alcohol Tester
	Use high precision and sensitive semiconductor sensor 2.0' lattice black and white screen with digital display 3 Test unit (%BAC, %BAC, mg/L) to be selected The numbner of Test may be display. Measurement range : 0.000~0.199% BAC; 0.000~1.990%BAC; 0.000~0.995mg/L (BrAC). Loud alarm sounds Accuracy : 0.010% BAC. Working temperature : 00C ~ 550C.

Mangal, AlcoTest	Professional Breath Alcohol Tester
	Build in DSP microporcessor technology. Warm-up time: less then 20 seconds Unique 4 digits LCD display Illumination button for use in dark Blowing time: 3 to 5 seconds continuously. Measurement Range: 0.000~0.200% BAC;0.000 ~2.000 BAC,0.000~1.000mg/L BrAC Accuracy: +/- 0.015%BCA at 0.05%BAC,+/- 0.05 mg/L at 0.25 mg/L (+/- 0.01 % at 0.05 %;+/- 0.10% at 0.50%) Display Range: 0.000-0.500mg/L BrAC (0.100% BAC,1.000%.BAC) Operating Temperature: -5 to % 0 degree C 9 Volt Battery.
Mangal, CA-2000	Professional Breath Alcohol Tester
	Warm-up Time below 20 seconds Response Time within 3 seconds Sensor Semiconductor oxide sensor Mouth Piece 5 x wasshable Mouthpiece Certified by the US food and drug administration (FDA) Adavanced sensor Indiation of BrAC: 0.00 to 2.00 mg/litre % BrAC Mouth Piece: 5 x Washable Mouthpiece 9 Volt Battery 